

## Section 5 Treatment and Recovery Areas

This part to be photocopied for the appropriate number of treatment and recovery areas in the practice.

<b>Practice/clinic name:</b>			
<b>Treatment/recovery area number:</b>			
<b>Type of area inspected:</b>	<input type="checkbox"/> Treatment	<input type="checkbox"/> Recovery	
<b>Type of treatment area:</b>	<input type="checkbox"/> IS	<input type="checkbox"/> IVS	<input type="checkbox"/> Other (please state below)
	Other:		

A. Facilities			Yes	No	Information Sources	Comments
1	A	There is access for emergency services to the area.....	<input type="checkbox"/>	<input type="checkbox"/>	Section 2.1, CSD	
2	A	There is adequate access for the clinical team around the chair to deal with medical emergencies and sedation-related complications.....	<input type="checkbox"/>	<input type="checkbox"/>		
3	A	There is good lighting in the area.....	<input type="checkbox"/>	<input type="checkbox"/>		

B. Unit/chair			Yes	No	Information Sources	Comments	
1	A	The dental chair can be rapidly moved to a head-down tilt position..... <b>N/A</b>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Domain F, SAAD Checklist	

C. Sedation equipment			Yes	No	Information Sources	Comments
		<b>NIBP monitor</b>			Section 2.1.2, CSD	
1	A	Non invasive blood pressure monitor.....				
2	A	Maintained in accordance with manufacturer's instructions.....				
		<b>Pulse oximeter</b> ..... N/A <input type="checkbox"/>				
3	A	Continuous pulse oximeter is available for use prior to and during treatment under sedation and has an audible alarm*.....				
4	A	There is a calibration and/or service history according to the manufacturer's instructions for the pulse oximeter.....				
5	A	Last service date of the pulse oximeter: (certificate required)..... <input type="text"/>				
6	A	<b>Inhalation sedation</b> ..... N/A <input type="checkbox"/>			Section 2.1.1, CSD	
7	A	The dedicated, purpose-designed IS machine has:				
8	A	• a minimum delivery of 30% oxygen.....				
9	A	• an emergency N2O cut off.....				
10	A	There is a calibration and service history according the manufacturer's instructions for the IS machine.....				
11	A	Last service date of the IS machine: (certificate required)..... <input type="text"/>				
12	A	<b>Gases</b> ..... N/A <input type="checkbox"/>			Section 2.1.1, CSD  Domain F, SAAD Checklist	
13	A	Gas cylinders are stored safely and securely according to current safety requirements.....				
14	A	Central gas supply system in date and serviced..... N/A <input type="checkbox"/>				
15	A	Last service date of the central gas supply: (certificate required)..... <input type="text"/>				
16	A	A record is kept of checks made by a suitably trained and qualified member of staff prior to each session.....				

\*Pulse oximeter with clearly visible display preferred.

C. Sedation equipment (continued)					Yes	No	Information Sources	Comments
17	A	Equipment for use with advanced techniques**	N/A				Domain F, SAAD Checklist	
18	A	ECG.....	N/A					
19	A	• Maintained, serviced and calibrated according to manufacturer's instructions.....						
20	A	Capnography.....	N/A					
21	A	• Maintained, serviced and calibrated according to manufacturer's instructions.....						
22	A	Syringe driver.....	N/A					
23	A	• Maintained, serviced and calibrated according to manufacturer's instructions.....						
24	A	Mucosal atomisation device.....	N/A					

**\*\*Currently NHS Scotland does not permit advanced sedation techniques in NHS General Dental Practice.**

D. Sedation drugs					Yes	No	Information Sources	Comments	
1	A	Midazolam (5mg/5ml strength*).....	N/A				Section 2.1.1, CSD		
2	A	Flumazenil (0.5mg/5ml) (if benzodiazepines are used)**.....	N/A						
3	A	Other drug(s).....	N/A						
		The sedation drugs recorded above are:							
4	A	• stored safely.....							
5	A	• in date - stocks checked regularly and checks logged.....							
6	A	Drug labels available for syringes.....	N/A						
7	A	Cannulae for IV access.....	N/A						
8	A	Supplemental oxygen is available and can be given, if required.	N/A						
9	A	Pocket masks with oxygen port***.....	N/A						

\* N.B. This is a lower concentration of midazolam than required for oromucosal administration. See Section 2D Item 10.  
 \*\* See Section 2D Item 14 and footnote \*\*. \*\*\* See Section 2D Item 19 and footnote \*\*\*.

